

Optik 1 : 3600/750  
 Optik 2 : 2400/750

Copying of this document, and giving it to others and the use of information contained herein, are forbidden without express authority. Customers are liable to the payment of damages. All rights are reserved in the case of the grant of a patent or the registration of a utility model or design.

Nr.	El.	Wellenl	T	Sbr	AL-BASE	HG-BASE	ZN-BASE			1	2	3	4	5
1	Cd	228.802	I	25			0.001-0.006			*				
2	Bg	231.450	0	75			REF. f. cd			*				
3	Si	251.612	I	50	0.001-3	0.001-0.1				*				
4	Zn	267.053	I	150			REFERENZ			*				
5	Hg	285.213	I	50	0.005-0.3		0.004-0.1			*				
6	Hg	291.552	I	75		REFERENZ				*	*			
7	Al	305.993	I	75	REFERENZ		3-29			*	*			
8	Be	313.042	II	50	0.0001-0.05 a	0.0001-0.01 a				*	*			
9	Sn	317.502	I	50	0.005-0.5		0.0004-0.01			*	*			
10	Cu	324.754	I	50	0.005-0.5	0.001-0.2	0.001-0.5			*	*			
11	Zn	334.502	I	50	0.005-5	0.005-1				*	*			
12	Zr	343.823	II	50	RESERVE					*	*			
13	Ni	352.454	I	50	0.001-0.6	0.0002-0.01 b	0.0005-0.04			*	*			
14	Fe	358.119	I	25	0.005-2 c	0.0005-0.01 b	0.001-0.12 c			*	*			
15	Fe	371.994	I	50	0.005-2		0.001-0.12			*	*			
16	Hg	382.935	I	50	0.3-10					*	*			
17	Si	390.553	I	50	3-15					*	*			
18	Al	394.403	I	50		8-10	1-3			*	*			
19	Mn	403.449	I	50	0.005-0.7	0.001-0.5				*	*			
20	Pb	405.782	I	25	0.001-0.1	0.003-0.01	0.001-0.01			*	*			
21	Cr	425.435	I	50	RESERVE					*	*			
22	Zn	481.053	I	150			REFERENZ			*	*			
23	Ti	498.173	I	50	RESERVE					*	*			
24	Cu	510.554	I	50	0.5-5		0.5-3			*	*			

- a.) Customer samples required
- b.) Diese Elemente mit SAFT II ausrüsten
- c.) Eichkurve an FRM, FRS

Alle Rechte vorbehalten. Nachdruck, Vervielfältigung und Verbreitung, auch auszugsweise, ist ohne schriftliche Genehmigung der Analytischen Instrumente GmbH. In der Bundesrepublik Deutschland ist die Nachdruck- und Vervielfältigung von geschützten Texten ohne schriftliche Genehmigung strafbar.

	Datum	Name
gez.:		<i>[Signature]</i>
gepr.:		
geö.:		

9/88/797  
 08.09.1997



Analytical Instruments GmbH  
 Boschstraße 10 · D-47533 Kleve  
 ☎ (02821) 892-0 · T x 811734 sptro d  
 Fax (02821) 23144